

26.09.2017

TAMILNADU MEDICAL SERVICES CORPORATION LTD.
TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF C-
ARM TO DEPT.OF SURGICAL GASTROENTEROLOGY, GOVT. RAJAJI HOSPITAL,
MADURAI UNDER PMSSY SCHEME

TENDER NO. 90/CARM/PMSSY-GRHM/TNMSC/ENGG/2017, Dt.22.02.2017

Corrigendum

a) The following corrigendum are issued:-

Sl. No.	Tender Document Reference	Instead of	Read as
1.	Page No.53 Section VI: Technical Specification And refer and corrigendum dated 31.08.2017 & 19.09.2017 1.C-Arm Type I Generator	80 KHz High frequency	80 KHz High frequency or more
2.	Page No.53 Section VI: Technical Specification And refer and corrigendum dated 31.08.2017 & 19.09.2017 1.C-Arm Type I X-ray tube	Housing heat capacity:1,900,000,HU	Housing heat capacity:1,900,000,HU or more
3.	Page No.53 Section VI: Technical Specification And refer and corrigendum dated 31.08.2017 & 19.09.2017 1.C-Arm Type I Hard Copy option	Image capture Device	Image capture System Should have in built 1 TB Hard Disk drive to store images
4.	Page No.53 Section VI: Technical Specification And refer and corrigendum dated 31.08.2017 & 19.09.2017		Add to the end of specification:- Image capture System Should have in built 1 TB

Sl. No.	Tender Document Reference	Instead of	Read as
	1.C-Arm Type II		Hard Disk drive to store images
5.	Page No.53 Section VI: Technical Specification And refer and corrigendum dated 31.08.2017 & 19.09.2017 1.C-Arm Type II	All functions remotely controlled from the C-arm control panel	Collimator and parallel shutters are controlled from C-arm panel
6.	Page No.59 Section VI: Bid form and Price Schedule And refer and corrigendum dated 31.08.2017 & 19.09.2017	Existing Text	Revised Text at Annexure

The due date is extended as follows :-

Sale of bidding document : up to 03.10.2017

Last date and time for receipt of bids : 04.10.2017, 3.00 PM

Date and time of opening of bids : 04.10.2017, 3.30 PM

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

Sd/-
General Manager (E)

ANNEXURE

PRICE SCHEDULE – REVISED

1	2	3	4	5					6	7	8
				PRICE FOR EACH UNIT							
Sch. No.	Item Description	Country of origin	Quantity & Unit	Ex-factory Ex-warehouse Ex-showroom off-the shelf for CIF (a)	Packing & forwarding (b)	Inland transport, Insurance and Incidental costs incidental to delivery (c)	Incidental services as listed in clause 7 of SCC (d)	Customs duty (e)	Unit price a+b+c+d+e	Total price per schedule for delivery at final destination (4 x 6)	GST – IGST/CGST/ SGST payable if contract is awarded
1	C-ARM (Type-I) as per specification		5 nos.								
1a	Difference in price between 9" image intensifier and 12" image intensifier		1 no.								
2	C-ARM (Type-II) as per specification		5 nos.								
2a.	Difference in price between 9" image intensifier and 12" image intensifier		1 no.								

i. Unit price in (6) (Rs in words)

1. C-ARM (Type I) as per specification

2. C-ARM (Type I) as per specification

a) Annual Maintenance Charges (labour only) per year for 5 years after 5 years free warranty maintenance period

1. C-ARM (Type I) as per specification

2. C-ARM (Type I) as per specification

a) Annual Maintenance Charges (Comprehensive) for 5 years / per year after 5 years free warranty maintenance period

1. C-ARM (Type I) as per specification

2. C-ARM (Type I) as per specification

Note:

- (a) In case of discrepancy between unit price and total price, the unit price shall prevail.
- (b) **This price schedule should be placed in separate sealed cover “Cover B”**
- (c) **GST tax applicable for Annual Maintenance Charges shall be indicated separately.**
- (d) **The bidder should indicate the HSN code of the equipment/ service and applicable GST rates.**
- (e) **The bidder may quote the prices for any one or both schedule separately (1 & 1a and 2 & 2a). However, the purchaser reserves the rights to order to any one of the schedule or combination of both, at their full discretion based on prices, utility factors and end user requirement.**

Place :

Signature of Bidder.....

Date :

Name

Business Address