

24.10.2017

**TAMILNADU MEDICAL SERVICES CORPORATION LTD.,**  
**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF**  
**PULSE OXIMETER TO DEPT. OF ANESTHESIA AT GOVT. RAJAJI HOSPITAL,**  
**MADURAI UNDER PMSSY SCHEME**  
**TENDER NO.174/MON/PMSSY-GRHM/TNMSC/ENGG/2017, DT.01.08.2017**

**Corrigendum**

a) The following corrigendum are furnished:-

Sl. No.	Tender Document Reference	Instead of	Read as
1.	Page No.55 Section VI: Technical Specification for Pulse Oximeter	15. Should be supplied with a reusable patient extension cable, reusable finger sensor and reusable wrap sensor.	15. Should be supplied: Reusable extension cable – 1 no. (Suitable for all three types of sensor) Reusable finger sensor – 1 no. each for Neonates, Pediatric and Adult Reusable wrap sensor – 1 no. each for Neonates, Pediatric and Adult.

b) The following clarification are furnished:-

Sl. No.	Tender Document Reference	Point Raised	Clarification furnished
2.	Page No.54 Section VI: Technical Specification for Pulse Oximeter <u>3. Should operate on 240 Ac/50Hz AC mains supply with built in rechargeable battery backup for a minimum of 16 hours.</u>	Request to amend as the battery backup atleast 6-8 hours.  Request to amend as the battery backup atleast 5-6 hours.	No Change. Hence, Published Specification Prevails.
3.	Page No.54 Section VI: Technical	Request to amend as the pulse rate range 30 to 240 or	No Change. Hence, Published Specification

Sl. No.	Tender Document Reference	Point Raised	Clarification furnished
	<p>Specification Specification for Pulse Oximeter 5. Should measure and display pulse rate and SPO2% in the range: <u>Oxygen Saturation: 0-100%</u> <u>Pulse rate range: 18-321 BPM</u></p>	250	Prevails.
4.	<p>Page No.54 Section VI: Technical Specification Specification for Pulse Oximeter <u>7. Should have a internal memory capacity of 70 hours to store patient in the internal memory to access the patient information using the optional data management software.</u></p>	Request to amend as the trend memory of 70 hrs download software is considered as optional.	No Change. It is not indicated as optional feature. This is a standard requirement only. Hence, Published Specification Prevails.
5.	<p>Page No.54 Section VI: Technical Specification Specification for Pulse Oximeter <u>8. Should have their own proprietary pulse oximetry standard and should work in low perfusion and motion artifact conditions as well. Should submit clinical paperwork about the oximetry technology used in the product.</u></p>	<p>Request to amend as the perfusion index and pleth variability index parameter with spO2 measurement</p> <p>Request to amend as the perfusion index (0.02 to 20) and fluid responsive pleth variability index (1 to 100%) with spO2 measurement and PR</p> <p>Request to amend as the perfusion index and pleth variability index can be added to enhance the patient outcome and monitor the fluid</p>	No Change. Hence, Published Specification Prevails.

Sl. No.	Tender Document Reference	Point Raised	Clarification furnished
		management.	
6.	Page No.55 Section VI: Technical Specification Specification for Pulse Oximeter <u>14. Should have patient security mode, recall function and user defined defaults.</u>	Request to clarify patient sensitive mode. Request to clarify patient safety mode.	No Change. Hence, Published Specification Prevails.
7.	Page No.55 Section VI: Technical Specification Specification for Pulse Oximeter <b>General Conditions</b> <u>5. All consumables required for installation, standardization and smooth functioning of equipment for 3 months period should be given free of cost along with supplied Equipments(if any)</u>	Request to indicate the consumables required with quantity.	It is indicated that consumables if any required should be supplied. If no consumable is required, it need not be supplied.
8.		Request to add the selectable pulse averaging mode with 2,4,8,10,12 sec.	

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-**

**Tender Inviting Authority**