

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,
TENDER FOR FIXING RATE CONTRACT FOR SUPPLY OF VIDEO NASAL
ENDOSCOPE TO VARIOUS INSTITUTIONS
TENDER NO.189/SCOPE/NRHM/TNMSC/ENGG/2017, DT.11.09.2017

Corrigendum

a) The following clarification are furnished:-

Sl. No.	Tender document reference	Point Raised	Clarification Furnished
1.	Page No.55 Section VI: Technical Specification Specification for Video Nasal Endoscope <u>150watts LED light source.</u>	Request to amend as 180 watts LED light source or more. It should have facility two LED output source with manual as well automatic light intensity control knob.	No Change. Hence, Published Specification Prevails.
2.	Page No.55 Section VI: Technical Specification Specification for Video Nasal Endoscope <u>Video Output: Composite signal to BNC socket, Y/C signal to S-VHS socket (2x)</u>	Request to amend as the Video output: USB, Composite Signal to BNC socket, Y/C signal to S-VHS socket (2x).	No Change. Hence, Published Specification Prevails.
3.	Page No.55 Section VI: Technical Specification Specification for Video Nasal Endoscope <u>It should be provided with automatic white balance with memory function.</u>	Request to amend as should be provided with automatic/manual white balance with memory function.	No Change. Hence, Published Specification Prevails.
4.	Page No.56 Section VI: Technical Specification Specification for Video Nasal Endoscope <u>The camera system should be provided with 2 programmable</u>	Request to amend as the camera system should be supplied with patient management software to retrieve old data. The recording system should have memory of	No Change. Hence, Published Specification Prevails.

Sl. No.	Tender document reference	Point Raised	Clarification Furnished
	<u>function keys on the camera head for control of camera functions, video printer/ recorder. The camera system should have image freeze control from the camera head.</u>	more 1 TB with software to record video and still images.	
5.	Page No.56 Section VI: Technical Specification Specification for Video Nasal Endoscope <u>All the instruments and equipments quoted should be manufactured by the same company excluding Trolley and TV.</u>	Request to amend as the bidding company (supplying company) should take the responsibility of After-sales service for all equipments and instruments supplied by them. Also should be supplying the instruments according to the quality norms and compatible to the other equipments supplied by them.	No Change. Hence, Published Specification Prevails.
6.	Page No.55 & 56 Section VI: Technical Specification Specification for Video Nasal Endoscope	Request to review and revise the technical specification.	No Change. Hence, Published Specification Prevails.

The following due date is extended:-

Sale of bidding document : up to 09.01.2018
 Last date and time for receipt of bids : 10.01.2018, 11.00 AM
 Date and time of opening of bids : 10.01.2018, 12.00 Noon

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

Sd/-
General Manager (E)