

17.01.2018

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF
REFRIGERATED CENTRIFUGE AND APHERESIS UNIT TO DEPT. OF
TRANSFUSION MEDICINE, GOVT. RAJAJI HOSPITAL, MADURAI UNDER PMSSY
SCHEME**

TENDER NO.167/BBK/PMSSY-GRHM/TNMSC/ENGG/2017, DT.01.08.2017

Corrigendum

a) The following corrigendum are issued:-

| Sl. No. | Tender document reference | Instead | Read as |
|----------------|--|---|--|
| 1 | Page No.54 Section VI: Technical Specification 1. Specification for Refrigerated Centrifuge (Cryofuge) <u>And refer corrigendum issued on 16.11.2017</u> | 9. Capacity – 6X2000ml to accommodate either 6X2000 ml bottles or 16 triple or quadruple Top and bottom bags of capacity 450 to 500ml blood. | 9. Capacity – 6X2000ml to accommodate both 6X2000 ml bottles and 16 triple or quadruple Top and bottom bags of capacity 450 to 500ml blood with wind shield rotor. |

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

Sd/-

General Manager (E)