

11.06.2018

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF
MULTICHANNEL MONITOR TO VARIOUS PRISON HOSPITAL AND AMBULANCE**

TENDER NO.290/MON/PRISON/TNMSC/ENGG/2018, DT.22.03.2018

Corrigendum

a) The following corrigendum are issued:-

Sl. No.	Tender Document Reference	Instead of	Read as
1.	Page No.54 Section VI: Technical Specification 1. Specification for Multichannel Monitor	Should be a semi modular type Multichannel monitor with minimum 15 inch color TFT display with Touch screen facility.	Should be a semi modular type Multichannel monitor with minimum 12 inch color TFT display with Touch screen facility.
2.	Page No.54 Section VI: Technical Specification 1. Specification for Multichannel Monitor	Should be upgradable with Thermal Array Printer, Dual IBP and Side stream based EtCo2 functions in future. Bidders to quote for the upgrades and modules optionally.	Should quote separately for the following upgrades and modules optionally in the price bid:- 1. Thermal Array Printer 2. Dual IBP 3. Side stream based EtCo2 functions. 4. Transport Monitor Module 5. 10 Lead ECG Cable
3.	Page No.61 Section VII: Bid Form and Price Schedule Price Schedule	Existing Text	Revised Text at Annexure - I

b) The following clarification are furnished:-

Sl. No.	Tender Document Reference	Point Raised	Clarification furnished
1.	Page No.54 Section VI: Technical Specification	Request to amend as trend facility for minimum 120 hours of graphical and tabular trends.	No Change. Hence, published specification prevails.

Sl. No.	Tender Document Reference	Point Raised	Clarification furnished
	<p>1. Specification for Multichannel Monitor</p> <p><u>Should have trend facility for 72 hours</u></p>		
2.	<p>Page No.54</p> <p>Section VI: Technical Specification</p> <p>1. Specification for Multichannel Monitor</p>	<p>Request to add the following technical points:-</p> <p>a) Should have facility to detect sudden critical blood pressure changes occur between periodic NIBP measurements to a precise continuous monitoring.</p> <p>b) Should have 12 lead ECG simultaneous display using 10 Lead ECG cable as on optional facility. Price should be quoted separately.</p> <p>c) Should have facility to update to interface the ventilator and Aneasthesia machine to view the parameters in the monitor.</p> <p>d) Should have venous puncture mode.</p> <p>e) Storage of data in monitor for minimum of 72 hours. Full disclosures, Alarm data, etc. for prison environment for any of the review purposes.</p> <p>f) Should have facility to print all the parameters and data using a direct connective printer.</p> <p>g) Should have upgradeable facility for the remote viewing</p>	<p>No Change. Hence, published specification prevails.</p>

Sl. No.	Tender Document Reference	Point Raised	Clarification furnished
		capability to monitor data from a distance even when the clinical is away from the hospital premises, through mobile phone of the clinician.	

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

Sd/-
General Manager (E)

PRICE SCHEDULE - REVISED

1	2	3	4	5					6	7	8
				PRICE FOR EACH UNIT							
Sch No.	Item Description	Country of origin	Quantity & Unit	Ex-factory Ex-warehouse Ex-showroom off-the shelf for CIF (a)	Packing & forwarding (b)	Inland transport, Insurance and Incidental costs incidental to delivery (c)	Incidental services as listed in clause 7 of SCC (d)	Customs duty (e)	Unit price a+b+c+d+e	Total price per schedule for delivery at final destination (4 x 6)	GST – IGST/CGST/ SGST payable if contract is awarded
1.	Multi Channel Monitor as per specification		50 Nos.								
	Optional Upgrades and Modules										
1a.	Thermal Array Printer		1 No.								
1b.	Dual IBP		1 No.								
1c.	Side stream based EtCo2 functions		1 No.								
1d.	Transport Monitor Module		1 No.								
1e.	10 Lead ECG Cable		1 No.								

i. Unit price in (6) (Rs. in words)

1. Multi Channel Monitor as per specification with HSN Code/rate of GST.....

Optional Upgrades and Modules

1a. Thermal Array Printer.....

1b. Dual IBP.....

1c. Side stream based EtCo2 functions.....

1d. Transport Monitor Module.....

1e. 10 Lead ECG Cable.....

a) Annual Maintenance Charges (labour only) per year /per unit for 7 years after 3 years free warranty maintenance period

1. Multi Channel Monitor as per specification with HSN Code/rate of GST.....

b) Annual Maintenance Charges (Comprehensive) for 7 years / per year / per unit after free warranty maintenance period

1. Multi Channel Monitor as per specification with HSN Code/rate of GST.....

Note:

- (a) In case of discrepancy between unit price and total price, the unit price shall prevail.
- (b) **This price schedule should be placed in separate sealed cover “Cover B”**
- (c) **GST applicable for Annual Maintenance Charges shall be indicated separately.**
- (d) **The bidder should indicate the HSN code of the equipment/ service and applicable GST rates.**
- (e) **The bidder should quote for the prices separately from column 5(a) to 5(e) and should not state “as inclusive”.**
- (f) **The bidder should indicate the prices for all optional items separately in the price schedule without including the same to the main price. But the prices of the optional items will not be taken into account for price evaluation. However, the purchaser reserves their rights to finalize these items at their sole discretion.**

Place :

Signature of Bidder.....

Date :

Name

Business Address