

25.09.2018

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**TENDER FOR SUPPLY AND INSTALLATION OF FULLY AUTOMATED
COAGULATION ANALYZER TO DEPT. OF PATHOLOGY, RAJIV GANDHI GOVT.**

GENERAL HOSPITAL, CHENNAI

TENDER NO.351/ANL/RGGGH/TNMSC/ENGG/2018, DT.27.07.2018

Corrigendum

a) The following clarification are furnished:-

Sl. No.	Tender document reference	Point raised	Clarification furnished
1.	Page No.54 Section VI: Technical Specification 1. Specification for Fully Automated Coagulation Analyzer 3. <u>It should have a throughput at least 35 PT/Hr and 15 APTT/Hr.</u>	Request to amend as should have a throughput at least 100 PT/Hr and 100 APTT/Hr.	No Change. Hence, the published specification prevails.
2.	Page No.54 Section VI: Technical Specification 1. Specification for Fully Automated Coagulation Analyzer 4. <u>It should have only the principle of scattered light detection method for clotting, colorimeter for chromogenic assay and later enhanced turbidimetric for D Dimer.II</u>	Request to amend as should have the principle of scattered light detection/Mechanical method for clotting, colorimeter for chromogenic assay and later enhance turbidimetric.	No Change. Optical light scatter is only required. Hence, the published specification prevails.
3.	Page No.54 Section VI: Technical Specification	Request to amend as should use minimum 4 wavelengths of LEDs, between 340 nm to 705nm for	No Change. 4 Wavelengths are not needed. Hence, the

Sl. No.	Tender document reference	Point raised	Clarification furnished
	<p>1. Specification for Fully Automated Coagulation Analyzer</p> <p>5. <u>It should use 3 wavelengths of LEDs, 640nm, 405nm and 575 nm for testing.</u></p>	testing.	published specification prevails.
4.	<p>Page No.54</p> <p>Section VI: Technical Specification</p> <p>1. Specification for Fully Automated Coagulation Analyzer</p> <p>8. <u>It should have minimum 3 sample racks and should take atleast 26 samples.</u></p>	Request to amend as should have minimum 5 sample racks and should take at least 50 tubes.	No Change. Hence, the published specification prevails.
5.	<p>Page No.54</p> <p>Section VI: Technical Specification</p> <p>1. Specification for Fully Automated Coagulation Analyzer</p> <p>17. <u>It should have and inbuilt color touch screen LCD facility for easy operator intervention and understanding.</u></p>	Request to amend as should have and inbuilt/external color touch screen LCD facility for easy operator intervention and understanding.	No Change. Hence, the published specification prevails.
6.	<p>Page No.54</p> <p>Section VI: Technical Specification</p> <p>1. Specification for Fully Automated Coagulation Analyzer</p> <p>18. <u>It should work on windows</u></p>	Request to amend as should work on windows software.	No Change. Hence, the published specification prevails.

Sl. No.	Tender document reference	Point raised	Clarification furnished
	<u>software and should not use a separate computer.</u>		
7.	Page No.55 Section VI: Technical Specification 1. Specification for Fully Automated Coagulation Analyzer <u>22. It should have in-built thermal graphic printer.</u>	Request to amend as should have in-built thermal graphic printer or external printer.	No Change. Hence, the published specification prevails.
8.	Page No.55 Section VI: Technical Specification 1. Specification for Fully Automated Coagulation Analyzer <u>28. It should use only single reaction cuvettes only and not any rotor or multi sample cassette/plates or cuvettes roll.</u>	Request to amend as should use only single reaction cuvettes during testing to minimize the wastage of cuvettes.	No Change. Hence, the published specification prevails.

The following due date is extended:-

Sale of bidding document : up to 03.10.2018

Last date and time for receipt of bids : 04.10.2018, 03.00 PM

Date and time of opening of bids : 04.10.2018, 04.00 PM

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-
General Manager (E)**