

30.04.2019

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,
PROCUREMENT OF COAGULATION ANALYZER TO GOVT. INSTITUTE OF
REHABILITATION MEDICINE, K.K. NAGAR, CHENNAI
LIMITED TENDER NO.353/LT/ANL/IRM/TNMSC/ENGG/2019, DT.21.01.2019
CORRIGENDUM

a) The following corrigendum are issued:-

Sl. No.	Tender document reference	Instead	Read as
1.	Page No.1 Annexure I Technical Specification	Existing Text	Revised Text at Annexure - I

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

Sd/-
General Manager (E)

1. SPECIFICATION FOR COAGULATION ANALYZER - REVISED

1. It should be basic single channel coagulation analyzer.
2. It should perform tests like PT INR, APTT.
3. It should have a inbuilt printer.
4. It should be a measuring principle of photometric.
5. Should have Incubation block with timer.
6. Should be supplied with 1000 nos. reaction cuvettes.
7. It should be an open system and able to use any brand reagents.
8. It should display the results in second, INR, Ratio.
9. It should be CE, ISO certified.
10. Mixing of the sample and reagent using Magnetic stirrer.