

**01.07.2019**

**TAMILNADU MEDICAL SERVICES CORPORATION LTD.,**  
**PROCUREMENT OF COAGULATION ANALYZER TO GOVT. INSTITUTE OF**  
**REHABILITATION MEDICINE, K.K. NAGAR, CHENNAI**  
**LIMITED TENDER NO.353/LT/ANL/IRM/TNMSC/ENGG/2019, DT.21.01.2019**  
**CORRIGENDUM**

a) The following corrigendum are issued:-

<b>Sl. No.</b>	<b>Tender document reference</b>	<b>Instead</b>	<b>Read as</b>
1.	Page No.1 Annexure I Technical Specification 1. Specification for Coagulation Analyzer <b>And refer corrigendum issued on 30.04.2019</b>	8. It should display the results in second, INR, Ratio.	8. It should display and print the results in second, INR, Ratio.
2.	Page No.1 Annexure I Technical Specification 1. Specification for Coagulation Analyzer <b>And refer corrigendum issued on 30.04.2019</b>	10. Mixing of the sample and reagent using Magnetic stirrer.	10. Mixing sample and reagents using magnetic stirrer/mixing of reagents using magnetic stirrer.

The Following due dates are extended:

Last Date and time for receipt of quotation : 05.07.2019 at 03.00 PM

Time and date Opening of quotation : 05.07.2019 at 03.30 PM

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-**  
**General Manager (E)**