

16.07.2019

**TAMILNADU MEDICAL SERVICES CORPORATION LTD.,**

**QUOTATIONS FOR SUPPLY OF MULTIPARA MONITOR WITH ECF, SPO2, NIBP,  
ETCO2 TO TRAUMA CARE CENTRE AT GOVT. HEADQUARTERS HOSPITAL,  
KANCHEEPURAM**

**TENDER NO.RSP15/NS/TNRSP-II/TNMSC/ENGG/2019, DT.30.04.2019**

**Corrigendum**

a) The following corrigendum are issued:-

<b>Sl. No.</b>	<b>Limited Tender document reference</b>	<b>Instead</b>	<b>Read as</b>
1.	Page no. 7 Technical Specification Specification for Multipara Monitor wit ECF, SPO2, NIBP, ETCO2 <b>9. Display</b>	d) Touch screen. Mouse navigation desirable	d) Touch screen.

b) The following clarification are furnished:-

<b>Sl. No.</b>	<b>Limited Tender document reference</b>	<b>Point raised</b>	<b>Clarification furnished</b>
1.	Page no. 7 Technical Specification Specification for Multipara Monitor wit ECF, SPO2, NIBP, ETCO2 <b>10. Data</b> b. <u>Should be able to connect with telemetry transceiver.</u>	Request to delete this point.	No Change. Hence, published specification prevails.
2.	Page no. 8 Technical Specification Specification for Multipara Monitor wit ECF, SPO2, NIBP, ETCO2	Request to delete this point.	No Change. Hence, published specification prevails.

Sl. No.	Limited Tender document reference	Point raised	Clarification furnished
	<p><b>10. Data</b></p> <p>e) <u>Standby/back-up GSM SIM card in case of failure of one network. Wherever network are robust, will work on a single GSM SIM.</u></p>		
3.	<p>Page no. 8 Technical Specification Specification for Multipara Monitor wit ECF, SPO2, NIBP, ETCO2</p> <p><b>10. Data</b></p> <p>f. <u>Data transfer should be enabled under Wi-Fi/2G/3G to the identified destination servers</u></p>	Request to delete this point.	No Change. Hence, published specification prevails.
4.	<p>Page no. 8 Technical Specification Specification for Multipara Monitor wit ECF, SPO2, NIBP, ETCO2</p> <p><b>10. Data</b></p> <p>g. <u>Auto reconnect facility, whenever network is absent.</u></p>	Request to delete this point.	No Change. Hence, published specification prevails.
5.	<p>Page no. 8 Technical Specification Specification for Multipara Monitor wit ECF, SPO2, NIBP, ETCO2</p> <p><b>10. Data</b></p> <p>h. <u>Intelligent Queue</u></p>	Request to delete this point.	No Change. Hence, published specification prevails.

Sl. No.	Limited Tender document reference	Point raised	Clarification furnished
	<u>management.</u> <u>i. Noise removal algorithm to be built-in to give accurate ECG in a running ambulance on roads which are reasonably bumpy. Ambulance need not stop while capturing the ECGs</u>		
6.	Page no. 8 Technical Specification Specification for Multipara Monitor wit ECF, SPO2, NIBP, ETCO2 <b>11. ECG</b> <u>h. ECG to be transmitted in at fixed intervals of every 5 minutes, with provision to send the specific desktops and/or PDA phones.</u> <u>i. Facility to forward and distribute ECG &amp; ST map form the back-end should be available.</u> <u>j. Option of activation of transmission at any point of time.</u> <u>k. Live streaming option to be made available, if required and should be quoted separately</u>	Request to delete this point.	No Change. Hence, published specification prevails.

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-  
General Manager (E)**