

07.08.2019

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**QUOTATIONS FOR SUPPLY OF LAB AUTOMATIC BLOOD GAS ANALYZER, ABG
MACHINE HAND HELD ANALYZER AND AUTOMATIC BIO-ANALYZER TO
TRAUMA CARE CENTRE AT GOVT. HEADQUARTERS HOSPITAL,
KANCHEEPURAM**

TENDER NO.RSP12/NS/TNRSP-II/TNMSC/ENGG/2019, DT.30.04.2019

Corrigendum

a) The following corrigendum are issued:-

Sl. No.	Tender document reference	Instead	Read as
1.	Page no.11 Annexure – I : Technical Specification 3. Specification for Automatic Bio Analyzer And refer corrigendum issued on 01.08.2019	14. Reaction Cuvette: permanent hard glass cuvettes or disposal cuvettes. If disposable cuvette is offered, the unit rate per cuvette shall be quoted separately.	14. Reaction Cuvette: minimum 72 permanent hard glass cuvettes maintained at 37 degree \pm 2 degree or disposal cuvettes. If disposable cuvette is offered, the unit rate per cuvette shall be quoted separately.
2.	Page no.11 Annexure – I : Technical Specification 3. Specification for Automatic Bio Analyzer And refer corrigendum issued on 01.08.2019	19. The water consumption should be minimum. Specify the maximum throughput.	19. The water consumption should be minimum. Specify the maximum water consumption of the analyzer

The Following due dates are extended:

Last Date and time for receipt of quotation : 13.08.2019 at 03.30 PM

Time and date Opening of quotation : 13.08.2019 at 04.00 PM

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-
General Manager (E)**