

07.08.2019

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,
TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF
DEFIBRILLATOR WITH EXTERNAL PACEMAKER TO VARIOUS GOVERNMENT
MEDICAL INSTITUTIONS
TENDER NO.493/DEFIB/RC/TNMSC/ENGG/2019, DT.14.06.2019
CORRIGENDUM

a) The following corrigendum are issued:-

Sl. No.	Tender document reference	Instead of	Read as
1.	Page No.54 Section VI: Technical Specification Specification for Defibrillator with External Pacemaker	1. The defibrillator should have biphasic technology having energy selection of 2-200 joules.	1. The defibrillator should have biphasic technology having energy selection of 2-300 joules.
2.	Page No.54 Section VI: Technical Specification Specification for Defibrillator with External Pacemaker	3. Machine should have more than 8" TFT screen.	3. Machine should have more than 7" TFT screen.
3.	Page No.54 Section VI: Technical Specification Specification for Defibrillator with External Pacemaker		Add the following point at end of the technical specification:- 17. The defibrillator should have Real time CPR Help with displaying compression depth & rate. 18. The pacemaker spike detector and display a mark on the ECG traces.

b) The following clarification are furnished:-

Sl. No.	Tender document reference	Point raised	Clarification Furnished
1.	Page No.54 Section VI: Technical Specification Specification for Defibrillator with External Pacemaker	Request to add the Rectilinear Biphasic technology.	Not required.

b) The following due date is extended:-

Sale of bidding document : up to 13.08.2019

Last date and time for receipt of bids : 14.08.2019, 03.00 PM

Date and time of opening of bids : 14.08.2019, 04.00 PM

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-
General Manager (E)**