

**13.08.2019**

**TAMILNADU MEDICAL SERVICES CORPORATION LTD.,**

**TENDER FOR SUPPLY AND INSTALLATION OF GAS CHROMATOGRAPHY WITH  
MASS SPECTROMETER TO MODERNIZATION OF WATER ANALYSIS**

**LABORATORY, KING INSTITUTE OF PREVENTIVE MEDICINE AND RESEARCH**

**INSTITUTE, GUINDY, CHENNAI**

**TENDER NO.500/CGY/KIPM/TNMSC/ENGG/2019, DT.11.07.2019**

**CORRIGENDUM**

a) The following corrigendum are issued:-

<b>Sl. No.</b>	<b>Tender document reference</b>	<b>Instead of</b>	<b>Read as</b>
1.	Page No.55 Section VI: Technical Specification 1. Specification for Gas Chromatograph with Mass Spectrometer		Add the following point at end of the technical specification:- 12. System: PC with latest processor and windows 7 or higher operating system, 8 GB RAM, ITB HDD, DVD, RW multi with 22" LED monitor, USB Key board, USB mouse along with all original operating system and mother board CDs. 13. Color printer. 14. 10µl syringe 15.5KVA online UPS 16. 1 hour backup with 26 AH SMF batteries. 17. Gas purification system 18. Helium gas cylinder with two stage regulator

**b) Commercial Points:-**

<b>Sl. No.</b>	<b>Tender document reference</b>	<b>Instead of</b>	<b>Read as</b>
1.	Page No.52 Section V: Schedule of Requirements	Delivery Schedule: 30 days from the date of purchase order.	Delivery Schedule: 60 days from the date of purchase order.
2.	Page No.57 Section VI A: Qualification Criteria 3. Common to both Manufacturer and Bidder	a) The bidder should have a manufacturing capacity of at least 18 Nos. similar capacity machines per annum certified by licensing authority/ chartered accountant.	a) The bidder should have a manufacturing capacity of at least 9 Nos. similar capacity machines per annum certified by licensing authority/ chartered accountant.

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-  
General Manager (E)**