

19.08.2019

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**TENDER FOR SUPPLY AND INSTALLATION OF AIR COOLED FREEZE DRYER
INDUSTRIAL TYPE TO INSTITUTE OF VETERINARY PREVENTIVE MEDICINE,
RANIPET**

**TENDER NO.503/FRD/AHD/TNMSC/ENGG/2019, DT.11.07.2019
CORRIGENDUM**

a) The following corrigendum are issued:-

Sl. No.	Tender document reference	Instead	Read as
1.	Page No.54 Section VI: Technical Specification Specification for Air Cooled Freeze Dryer (Industrial Type) Dry Chamber	The chamber should have an effective shelf load surface area of 1.30 m ² or more.	The chamber should have an effective shelf load surface area less than 1.30 m ² .
2.	Page No.55 Section VI: Technical Specification Specification for Air Cooled Freeze Dryer (Industrial Type) Dry Chamber	Drying chamber should be preferably in rectangle / square shape with suggestive number of 8 loading shelves for clear segregation of products and for easy loading.	Drying chamber should be preferably in rectangle / square shape with suggestive number of 8 loading shelves for clear segregation of products and for easy loading of trays with intershelf distance more than 100mm.
3.	Page No.60 Section VI: Technical Specification Specification for Air Cooled Freeze Dryer (Industrial Type)		Add the following point at end of the technical specification:- The freeze dryer machine should be installed in the first floor of Poultry Vaccine Section GMP Building of IVPM, Ranipet.

b) The following corrigendum are issued:-

Sl. No.	Tender document reference	Point raised	Clarification Furnished
1.	Page No.54 Section VI: Technical Specification Specification for Air Cooled Freeze Dryer (Industrial Type) Dry Chamber <u>Drying chamber should be preferably in rectangle / square shape with suggestive number of 8 loading shelves for clear segregation of products and for easy loading of trays with intershelf distance between more than 100mm.</u>	Request to clarify whether 10 loading shelves with shelf area of 1.42 m ² can be quoted.	8 loading shelves is sufficient but the surface area should not be more than 1.3 m ²
2.	Page No.13 Section II: Instruction to Bidders 10. Bid Currencies 10.1 <u>Prices shall be quoted in Indian Rupees.</u>	Request to clarify whether price can quote in foreign currency instead of Indian rupee.	No Change. Hence, published terms prevails.

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-
General Manager (E)**