

09.09.2019

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF
OT LIGHT – MOBILE, SHADOWLESS TYPE TO VARIOUS GOVT. INSTITUTIONS**

TENDER NO.511/OTL/RC/TNMSC/ENGG/2019, DT.31.07.2019

CORRIGENDUM

a) The following corrigendum are issued:-

Sl. No.	Tender document reference	Instead	Read as
1.	Page No.52 Section V: Schedule of Requirements	Existing Text	Revised Text at Annexure - I

b) The following clarification are furnished:-

Sl. No.	Tender document reference	Point raised	Clarification furnished
1.	Page No.54 Section V: Technical Specification 1. Specification for OT Light – Mobile, Shadowless Type B1. Technical Requirements <u>1. Should consist of a Single movable light head/ dome with removable autoclave sterilizable handler light source reflectors mirrors. Mechanism to adjust the focus. Heavy base of 24 inches diameter with castors.</u>	Request to amend as LED technology.	No Change. The given requirement is LED technology (Refer point no.B1(4))
2.	Page No.54 Section V: Technical Specification	Request to amend as more than 50000 Hrs.	No Change. The given requirement is minimum. Higher specification may

Sl. No.	Tender document reference	Point raised	Clarification furnished
	1. Specification for OT Light – Mobile, Shadowless Type B1. Technical Requirements <u>5. Life of light source: 20000 hours or more</u>		also be quoted.
3.	Page No.54 Section V: Technical Specification 1. Specification for OT Light – Mobile, Shadowless Type B1. Technical Requirements <u>7. Colour temperature: more than 3000 K</u>	Request to amend as variable colour temperature.	No Change. The given requirement is minimum. Higher specification may also be quoted.
4.	Page No.54 Section V: Technical Specification 1. Specification for OT Light – Mobile, Shadowless Type B1. Technical Requirements <u>8. Maximum illumination level: more than 90000 lux at 1 metre</u>	Request to amend as 16000 lux at metre	No Change. The given requirement is minimum. Higher specification may also be quoted.
5.	Page No.54 Section V: Technical Specification 1. Specification for OT Light – Mobile, Shadowless Type B1. Technical Requirements <u>10. Brightness control to allow full adjustment from zero to</u>	Request to amend as 50% to maximum illumination.	No Change. The given requirement is minimum. Higher specification may also be quoted.

Sl. No.	Tender document reference	Point raised	Clarification furnished
	<u>maximum illumination</u>		
6.	Page No.54 Section V: Technical Specification 1. Specification for OT Light – Mobile, Shadowless Type B1. Technical Requirements <u>11. Illumination backup through multiple bulb use or spare bulb activation.</u>	Request to delete this point.	No Change. Hence, published specification prevails.
7.	Page No.54 Section V: Technical Specification 1. Specification for OT Light – Mobile, Shadowless Type <u>(C) 4. Should conform to ISI/EUMDD/European CE or USFDA with necessary marking on the product.</u>	Request to amend as CE European and USFDA certification. Request to delete the European CE/USFDA and include as Indian CE.	No Change. Hence, published specification prevails.

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-
General Manager (E)**

**SECTION – V - REVISED
SCHEDULE OF REQUIREMENTS**

Sch. No.	Brief Description	Unit	Tentative Qty.	Bid security (Rs.)
1.	OT Light – Mobile, Shadowless Type as per specification	Nos.	15	Rs.12,000/-

Delivery Schedule: - 30 days from the date of purchase order for the full quantity and for increased quantity the delivery period will be increased proportionately.

Place of Delivery:- Delivery places will be intimated at the time of placement of order.

Note :

- 1) The rate contract should be valid for a period of 2 years from the date of acceptance of the tender and extendable for further period on mutual acceptance.
- 2) The quantity indicated is tentative and will vary as per the actual requirement during the rate contract period. The supplier should have no claim against the purchase on the actual quantity of orders placed during the rate contract period.
- 3) During the rate contract period, further quantity may be ordered for any other Govt. Medical Institutions at the same rate and terms & conditions.