

20.09.2019

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**TENDER FOR SUPPLY OF ELECTRIC DERMATOME AND SKIN GRAFT MESHER
TO VARIOUS GOVT. HOSPITALS**

**TENDER NO.467/BURNS/DME/TNMSC/ENGG/2019, DT.31.05.2019
CORRIGENDUM**

a) The following corrigendum are issued:-

Sl. No.	Tender document reference	Instead	Read as
1.	Page No.1,3 & 5	Tender for supply of Electric Dermatome and Skin Graft Mesher to Various Govt. Hospitals	Tender for supply of Battery Operated Dermatome and Skin Graft Mesher to Various Govt. Hospitals
2.	Page No.52 Section V: Schedule of Requirements	Existing Text	Revised Text at Annexure - I
3.	Page No.54 Section VI: Technical Specifications 1. Specification for Electric Dermatome	Existing Text	Revised Text at Annexure - II
4.	Page No.61 Section VII: Bid form and Price Schedule Price Schedule	Existing Text	Revised Text at Annexure - III

b) The following clarification are furnished:-

Sl. No.	Tender document reference	Point raised	Clarification Furnished
1.	Page No.54 Section VI: Technical Specification 1. Specification Battery Operated Dermatome <u>2. Should have thickness</u>	Request to amend as control adjustment ranges from 0 to 0.030 in (0.75mm) in 0.02 in (0.050mm) increments.	No Change. Hence, published specification prevails.

Sl. No.	Tender document reference	Point raised	Clarification Furnished
	<u>control adjustment ranges from 0 to 0.048 in (1.20 mm) in 0.004 in (0.10 mm) increments.</u>		

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-
General Manager (E)**

**SECTION – V - REVISED
SCHEDULE OF REQUIREMENTS**

Sch. No.	Brief Description	Unit	Qty.	Bid security (Rs.)
1.	Battery Operated Dermatome as per specification	Nos.	12	Rs.60,000/-
2.	Skin Graft Mesher as per specification	Nos.	12	Rs.22,000/-

Delivery Schedule: - 30 days from the date of purchase order.

Place of Delivery:- The above equipment should be delivered at

Sl. No.	Name of the Institution	Battery Operated Dermatome (Qty.)	Skin Graft Mesher (Qty.)
1	KAPV Viswanathan Govt. Medical College Hospital, Trichy	2	2
2	Govt. Medical College Hospital, Vellore	2	2
3	Govt. Medical College Hospital, Salem	2	2
4	Govt. Medical College Hospital, Coimbatore	2	2
5	Govt. Medical College Hospital, Thanjavur	2	2
6	Govt. Rajaji Hospital, Madurai	2	2
	Total	12	12

1. SPECIFICATION FOR BATTERY OPERATED DERMATOME - REVISED

It is a skin grafting instrument to provide variable graft thickness and width capabilities.

1. Should have a battery operated surgical skin grafting instrument.
2. Should have thickness control adjustment ranges from 0 to 0.048 in (1.20 mm) in 0.004 in (0.10 mm) increments.
3. Should have individual graft widths of 1 in., 1.5 in., 2 in., 3 in., and 4 in. (2.5 cm, 3.8cm, 5.1 cm, 7.6 cm, 10.2 cm) are obtained with five width plates.
4. Should have two stainless steel machine screws secure the plates to the underside of the instrument.
5. The plates are easily fastened and removed with the screwdriver provided.
6. The dermatome should have powered by an ironless rotor, low inertia motor, which provides nearly vibration – free power.
 - I. Maximum protection class: class 1 degree of protection against electrical shock. Type BF II.
 - II. Handpiece.
 - a. Vibration and shock: Standard commercial practice.
 - b. Nominal speed: 4,500 – 5,500 cycles/minute.
 - c. Nominal speed: 4,500-5,500 cycles/minute.
7. Should be battery operated dermatome (cordless dermatome). The battery should be rechargeable and should be supplied with suitable charger.

PRICE SCHEDULE - REVISED

1	2	3	4	5					6	7	8
				PRICE FOR EACH UNIT							
Sch No.	Item Description	Country of origin	Quantity & Unit	Ex-factory Ex-warehouse Ex-showroom off-the shelf for CIF (a)	Packing & forwarding (b)	Inland transport, Insurance and Incidental costs incidental to delivery (c)	Incidental services as listed in clause 7 of SCC (d)	Customs duty (e)	Unit price a+b+c+d+e	Total price per schedule for delivery at final destination (4 x 6)	GST – IGST/CGST/ SGST payable if contract is awarded
1.	Battery Operated Dermatome as per specification		12 Nos.								
2.	Skin Graft Mesher as per specification		12 Nos.								

i. Unit price in (6) (Rs. in words)

- Battery Operated Dermatome as per specification with HSN Code/rate of GST.....
- Skin Graft Mesher as per specification with HSN Code/rate of GST.....

a) Annual Maintenance Charges (labour only) per year /per unit for 7 years after 3 years free warranty maintenance period

- Battery Operated Dermatome as per specification with HSN Code/rate of GST.....
- Skin Graft Mesher as per specification with HSN Code/rate of GST.....

b) Annual Maintenance Charges (Comprehensive) for 7 years / per year / per unit after free warranty maintenance period

- Battery Operated Dermatome as per specification with HSN Code/rate of GST.....
- Skin Graft Mesher as per specification with HSN Code/rate of GST.....

Note:

- In case of discrepancy between unit price and total price, the unit price shall prevail.
- This price schedule should be placed in separate sealed cover “Cover B”**
- GST applicable for Annual Maintenance Charges shall be indicated separately.**
- The bidder should indicate the HSN code of the equipment/ service and applicable GST rates.**
- The bidder should quote for the prices separately from column 5(a) to 5(e) and should not state “as inclusive”.**

Place : Signature of Bidder.....

Date : Name

Business Address