

TAMIL NADU MEDICAL SERVICES CORPORATION LIMITED

(A Government of Tamilnadu Undertaking)

417, PANTHEON ROAD, EGMORE, CHENNAI – 600 008.

ISO 9001:2008 Certified Organisation

Phone:28191890, 28190259 Fax No.28190636

CIN: U85110TN1994SGC027939

Ref. No.358/LT/CEN/DMS-GHTM/TNMSC/ENGG/2019

04.02.2019

INVITATION FOR QUOTATIONS FOR SUPPLY OF EQUIPMENT
THIS INVITATION DULY SIGNED ON ALL PAGES SHOULD BE ATTACHED
WITH THE QUOTE

From

The General Manager (E),
TamilNadu Medical Services Corporation Limited,
417, Pantheon Road, Egmore,
Chennai – 600 008.

To

Dear Sirs,

Sub: Invitation for Quotations for Supply of Microliter Centrifuge and
Microfuge to Govt. Hospital of Thoracic Medicine, Tambaram
Medicine, Chennai - Reg.

1. You are invited to submit your most competitive quotation for the following
goods:-

| Sl. No. | Brief Description of the Goods | Qty in No. | Delivery Period | Place of Delivery |
|---------|--------------------------------|------------|---|---|
| 1. | Microliter Centrifuge | 1 | Within 30 days from the date of receipt of supply order | BSL III TB culture facility laboratory, Govt. Hospital of Thoracic Medicine, Tambaram Medicine, Chennai |
| 2. | Microfuge | 1 | | |

2. Bid Price

2.1 The contract shall be for the full quantity as described above.
Corrections, if any, shall be made by crossing out, initialing,
dating and re writing.

- 2.2 All duties, taxes payable by the contractor under the contract shall be considered in the quoted price indicating the breakup value.
 - 2.3
 - a) The bidder should furnish the CIF value of the imported component with Customs duty separately in the price bid in addition to the other breakup value packaging & Forwarding, Inland transport, Installation and commissioning and other incidental charges
 - b) The bidders are advised to take in to consideration the savings available due to abolition of CVD and SAD for imports now due to GST implementation, and consider the benefits out of it as a reduction, while arriving at their bid price.
 - c) Similarly, the benefits that could be availed out of High Sea Sales/ Sales in course of import should also be considered as a reduction while arriving at their bid price.
 - d) GST shall be indicated separately.
 - 2.4 The Prices should be quoted in Indian Rupees only.
3. Each bidder shall submit only one quotation.
- 3.1. Earnest money deposit:- Each quotation should be submitted with an EMD of Rs.1,000/- by means of DD drawn in favour of "Tamilnadu Medical Services Corporation Limited, Chennai payable at Chennai.
 - 3.2 Quotation received without EMD amount will be treated as non-responsive.
 - 3.3 The bidder should furnish the GST registration no. for supply and services and the code no. for the goods quoted.
 - 3.4 The GST registered bidders are only eligible to participate in the quotation.
 - 3.5 The bidder should furnish the details of Bank name, Branch name, Account no., IFSC code and a copy of cancelled cheque leaf.
 - 3.6 The validity of the Demand Draft / Banker's Cheque furnished for EMD should be not less than 15 days.

4. **Validity of Quotation:-**

Quotation shall remain valid for a period not less than **45** days after the deadline date specified for submission.

5. **Evaluation of Quotations:-**

5.1 The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which

(a) are properly signed; and

(b) conform to the terms and conditions, and specifications

(c) Not furnishing Catalogues, Point wise compliance statement of Technical specification, and producing the sample of the quoted item for demo within one week notice, if called for would be treated as a violation of the terms of the tender and the bids would be treated as non responsive.

5.2 The evaluation of the prices will be done by adding the comprehensive AMC charges at discounted rate of 8% per annum with the supply price. Firms who are not quoted for the comprehensive AMC would be considered as non responsive.

The Quotation would be evaluated separately for each item. GST will not be taken into account in evaluation.

6. **Award of contract:-**

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

6.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.

6.2 The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.

6.3 For delay in supply of the goods beyond the stipulated delivery period, liquidated damages at 0.5% per week or part thereof of

undelivered portion of the contract, subject to a maximum of 10% of the contract value is leviable.

6.4 You shall furnish performance security for 5% of contract value by means of demand draft drawn in favour of Managing Director Tamilnadu Medical Service Corporation Limited Payable at Chennai which will be returned after satisfactory completion of warranty period.

6.5 Payment shall be made within 30 days after delivery of the goods,
against submission of bills with certification from the consignee for satisfactory completion of supply.

6.6 The item supplied shall carry a warranty of atleast 3 years for equipment from the date of acceptance of the goods by the consignee.

6.7 Annual Maintenance Charges (Comprehensive) should be quoted for 7 years / per year after 3 years warranty maintenance period for equipment.

7. You are requested to submit the quotations in a sealed cover superscribed **“Quotations for Procurement of Microliter Centrifuge and Microfuge to Govt. Hospital of Thoracic Medicine, Tambaram Medicine Chennai”** and quotations should reach Tamil Nadu Medical Services (A Govt. of Tamilnadu Undertaking), 417, Pantheon Road, Egmore, Chennai – 600008 on or before **11.30 hours** on **20.02.2019**.

8. Quotation will be opened at **12.00 hours** on **20.02.2019** in the presence of bidders.

9. We look forward to receive your quotations and thank you for your interest in this project.

**Sd/-
General Manager (E)**

PRICE SCHEDULE

| 1 | 2 | 3 | 4 | 5 | | | | | 6 | 7 | 8 |
|---------|--|-------------------|-----------------|---|--------------------------|---|-------------------------|------------------|------------|--|---|
| | | | | PRICE FOR EACH UNIT | | | | | | | |
| Sch No. | Item Description | Country of origin | Quantity & Unit | Ex-factory Ex-warehouse Ex-showroom off-the shelf for CIF (a) | Packing & forwarding (b) | Inland transport, Insurance and Incidental costs incidental to delivery I | Incidental services (d) | Customs duty (e) | Unit price | Total price per schedule for delivery at final destination (4 x 6) | GST – IGST/CGST/SGST payable if contract is awarded |
| 1 | Microliter Centrifuge as per specification | | 1 No. | | | | | | a+b+c+d+e | | |
| 2 | Microfuge as per specification | | 1 No. | | | | | | | | |

i. Unit price in (6) (Rs. In words)

1. Microliter Centrifuge as per specification as per specification with HSN code/ rate of GST.....
2. Microfuge as per specification as per specification with HSN code/ rate of GST.....

a) Annual Maintenance Charges (labour only) per year /per unit for 7 years after 3 years free warranty maintenance period

1. Microliter Centrifuge as per specification as per specification with HSN code/ rate of GST.....
2. Microfuge as per specification as per specification with HSN code/ rate of GST.....

b) Annual Maintenance Charges (Comprehensive) for 7 years / per year / per unit after free warranty maintenance period

1. Microliter Centrifuge as per specification as per specification with HSN code/ rate of GST.....
2. Microfuge as per specification as per specification with HSN code/ rate of GST.....

Note:

- (a) **In case of discrepancy between unit price and total price, the unit price shall prevail.**
- (b) **GST applicable for Annual Maintenance Charges shall be indicated separately.**
- (c) **The bidder should indicate the HSN code of the equipment/ service and applicable GST rate**
- (d) **The bidder should quote for the prices separately from column 5(a) to 5(e) and should not state “as inclusive” of all rate.**

Place :
Date :

Signature of Bidder.....
Name
Business Address.....

TECHNICAL SPECIFICATION

1. SPECIFICATION FOR MICROLITER CENTRIFUGE

1. Robust metal housing; compact design with chemical-resistant (coated) housing.
2. Easy-to-clean, smooth rotor chamber that is resistant to acids, alkalines, disinfectants used in the laboratory.
3. Low access height (≤ 23 cm) and space-saving design (≤ 24 cm \times 32 cm; W \times D) Approx.
4. Microprocessor-controlled centrifuge.
5. Standard rotor with a capacity of at least 18 positions; aerosol-tight (chemical-resistant coated); exchangeable.
6. Maintenance-free motor.
7. Maximal relative centrifuge force: 15,000 G.
8. Automatic lid lock, starting with and during run of rotor.
9. Option: Automatic opening at the end of the run.
10. Emergency unlock for electricity blackout.
11. LCD display; protected; showing time and relative centrifugal force or speed in rcf or rpm.
12. Speed adjustable in 100 rpm steps.
13. If a keypad is used, it should be foil protected.
14. Timer for runs between 30 seconds and 30 minutes and an option for continuous operation for longer runs.
15. Short time operation by pressing a time button for short spin.
16. Adjustment of running time in steps of 30 seconds.
17. Short acceleration time to maximum rcf in ≤ 20
18. Short breaking time from maximum rcf in ≤ 20 seconds.
19. Noise level: ≤ 58 dbA

2. SPECIFICATION FOR MICROFUGE

1. Microfuge/minispin for 0.2-2ml tubes for short-rapid spin down of reagents.
2. Variable speed microfuge Maximum speed 13,500 rpm Timer and pulse start 12 tube capacity Compact and quiet Lid safety mechanism, antimicrobial protection

3. Should be compact microcentrifuge for centrifugation of up to 12 x 1.5ml or 2.2ml microfuge tubes or four strips of 8 x 0.2ml tubes at speeds of up to 13,500rpm.
4. The unit should be supplied with adaptors for use with the fixed angle rotor to enable 0.2ml tubes to be centrifuged.
5. The rotors should be installed without the use of tools. Centrifugation speed can be displayed as either rpm or rcf and the timer can be set for run times of between 1 and 30 minutes.
6. In addition, a pulse button enables a rapid spin down by accelerating the rotor to the preset rpm then promptly decelerating. For safety, the door should remain locked while the rotor is in motion.
7. Should be supplied with Fixed angle rotor for 12 x 1.5 or 2.2ml tubes